

A microscopic image of numerous rod-shaped bacteria, likely E. coli, in shades of blue and green. The bacteria are scattered across the frame, with some in sharp focus and others blurred in the background.

OGI3 BioReactor

User Manual 2025

Contents

Contents	2
Device Manufacture	4
Safety Instructions.....	4
1. Operating the device.....	5
2. Device Description	6
3. Base Unit Overview.....	9
3.1. Front Panel	9
3.2. Rear Panel.....	10
3.3. Birds-Eye-View.....	11
4. Handling the device	12
4.1. Powering on	12
4.2. USB drive.....	12
4.3. Culture volume	13
4.4. Cleaning the device and its modules	13
4.5. Device Main Menu	15
4.6. Quick Measure	16
5. Calibration of the device.....	18
5.1. Optical Density and Calibration	18
5.2. Calibrating your Device	19
5.3. Manual OD Calibration.....	20
6. Setting up an experiment.....	21
6.1. Starting an experiment from the device.....	21
6.2. During an Experiment.....	22
6.3. Data format and handling.....	23
6.4. Saving Settings.....	23
Accessory Modules.....	25
7. Liquid Control Module (P/N OGI3-1011).....	25
7.1. Connecting the Liquid Control module.....	25
7.2. Fitting the tubing.....	26
7.3. Cleaning the pump tubing.....	28
7.4. Preparing for a liquid-controlled experiment.....	30
7.5. Turbidostat Experiment Settings	32
7.6. Turbidostat Media Usage Prediction.....	33
7.7. During a Liquid Control experiment	33

7.8.	Data format and handling.....	34
7.9.	Chemostat Functionality	34
7.10.	pH Control.....	34
8.	Oxygen Module (P/N OGI3-1008).....	35
8.1.	Connecting the oxygen module	35
8.2.	Calibrating the Oxygen Modules	35
8.3.	Measuring dissolved oxygen in experiments	38
8.4.	Data format and handling.....	38
8.5.	Notes.....	39
9.	Fluorescence Module (P/N OGI3-1014).....	40
9.1.	Connecting the Fluorescence Module.....	40
9.2.	Description of the Fluorescence Module.....	40
9.3.	Using the Fluorescence Module	41
9.4.	Data Format and Handling	41
10.	pH Module (OGI P/N OGI3-1012)	42
10.1.	Connecting the pH Module	42
10.3.1.	Configuration of pH Buffers for pH Calibration.....	43
10.3.2.	Calibration of the pH Module	44
10.4.	Measure pH in Experiments.....	45
10.5.	pH Control.....	45
10.6.	Data Handling	46
	Disposal	47
	Appendix A - Technical Specification	48
	Appendix B: Troubleshooting	49
	Calibration Errors and Warnings	49
	Oxygen Spot Errors	49
	Version Control.....	51



This symbol indicates an important point to consider during device use.

Device Manufacture

Device is manufactured by:

ŌGI Bio Ltd

Glencorse Building

Pentlands Science Park

Bush Loan

Penicuik

EH26 0PZ

Safety Instructions

If the device is not used as indicated in this manual, its functioning and safety may be impaired.

- Please turn the device off before replacing and installing optional modules.
- Power Supply must be connected to supply with a protective earth connection.
- The optional Fluorescence Measurement module (OGI3-1014) contains a UV LED. This has been tested and assessed in operating mode as Risk Group 0. Please ensure module is firmly secured to the Bioreactor with the fixing screw provided BEFORE turning on the Bioreactor power.
- [Do not unscrew the base of the Bioreactor](#) or modules to open them. If you have any issues with the device, please contact our team at help@ogibio.co.uk
- Do not position the Bioreactor so that it is difficult to operate or disconnect the device. The mains supply socket should be easily accessible in order that it can be disconnected quickly in case of emergency.
- The heating elements have a built in protection for overheating and will pause any in progress experiment if overheating is detected.

1. Operating the device

If the device is not used as indicated in this manual, its functioning and safety may be impaired.

Technical specifications can be found in [Appendix A](#).

Using the device outwith the following operating parameters may result in suboptimal performance:

- Operating temperature: 15°C to 60°C (note that temperature control is limited to 50°C).
- Relative humidity: 30% - 70%.
- Maximum operating altitude: 2000 meters.

2. Device Description

The ŌGI Bio OGI3 BioReactor (P/N OGI3-1010) is a small-scale bench-top bioreactor primarily designed to automate the culturing of microbes while providing accurate and on-demand culture analytics hands-free. The device comprises four individual, independent, reactors. Single culture flasks are held in the central flask holders of each reactor and are equipped with a magnetic stir bar for mixing and oxygenation. The optical density (OD) of each culture is measured at regular intervals throughout the experiment. The device has integrated heating elements and can maintain each of the culture flasks at desired temperatures. Silicon bungs are used to close the flask, with plugs provided so that the flask may be closed to the lab environment no matter which modules are installed. Miniature air filters (KIT-0306) can be inserted into vent holes on the bung to provide sterile air exchange with the headspace of the flask.

Additional optional accessory modules can be installed to add capabilities to the device:

WiFi Connectivity

OGI3 comes with a WiFi antenna and the ability to connect to a WiFi network to retrieve date information for experimental setup. In early 2025 this functionality will be extended to setting up, monitoring, and controlling experiments via our online Web App.

Liquid Control Module OGI3-1011

The Liquid Control module can be used to maintain the cultures at the desired optical densities, or in chemostat mode to provide a constant flow of fresh media into the culture flasks. This module contains 8 pumps for liquid handling: 4 input pumps to add fresh media into the culture flasks and 4 output pumps to remove the excess volume from the culture flask. Bungs containing needles are provided for use with the Liquid Control Module. See [Section 7](#) for more information.

Dissolved Oxygen Module OGI3-1008

After calibration with reference samples, the Dissolved Oxygen module can be used to measure the amount of dissolved oxygen in the culture media at regular intervals during an experiment. This module requires culture flasks with oxygen sensor spots (Part number A200017). See [Section 8](#) for more information.

pH Module OGI3-1012

The pH module can be used to measure the pH of the culture. The module has 4 pH probes included, one for each reactor. Used in conjunction with the Liquid Control Module (OGI-A-1011) the pH of a culture can be maintained at a set point. See [Section 10](#) for more information.

Sparging Module: OGI3-1013 – COMING 2025

The Sparging module will be used to bubble air into the head space or directly into the culture through a sparger. The module contains 4 pumps connected to sparging needles for insertion into each culture.

Fluorescence Module: OGI3-1014

The fluorescence module can be used to measure fluorescence of a variety of fluorophores in the culture flasks. It is possible to measure up to 3 different fluorescent signals simultaneously per flask. The wavelengths available as standard are:

Excitation: 365nm, 470nm, 590nm.

Emission: 400-450nm, 450-500nm, 500-550nm, 550-600nm, 600-650nm, 650-700nm, 700-750nm, 750-800nm.

All experiment data for the OGI3 and any attached modules is saved to a USB drive and can be analysed after the experiment. See [Section 9](#) for more information.

N.B. Any information relating to liquid control, pH, oxygen measurements or fluorescence measurements can only be utilised if these modules are connected to the device. Refer to the relevant sections of the user manual for further information.

3. Base Unit Overview

3.1. Front Panel

The front of the OGI3 BioReactor is shown below and the key interface points are labelled.




Figure 1: Front view of the OGI3 BioReactor

3.2. Rear Panel




Figure 2: Rear view of the OGI3 BioReactor

On the rear panel of the device:

- On/off switch
- Fuse
- Power supply port where the power supply is to be connected
- Liquid Control module connector.
- Micro USB ports for software updates to main BioReactor and Wi-Fi functionality

N.B. Please do not attempt to connect to WiFi Prog. unless to update the device software using one of our update packages. Board Prog. may be used in conjunction with provided script to view ODs live. Incorrect use of these ports could lead to damage to the device.

3.3. Birds-Eye-View




Figure 3: Overview of the complete BioReactor

From above, the locations of key points of interface with the device are highlighted. Note that the modules (Liquid Control, Oxygen, Fluorescence, pH, and Sparging) do not come as default with the device. Temperature control is included as standard.

4. Handling the device

4.1. Powering on

The socket for the power supply cable sits at the back of the OGI3 as shown in Figure 2. Ensure that the USB stick, temperature probes, and any optional modules/probes are connected to the device prior to powering on.

The device is switched on via the On/Off Switch on the rear of the Base Unit (Figure 2).

Upon turning on the device it will automatically detect which modules are installed, and on which reactors they are present. It will display a short message for each module it finds.

The device will then attempt to connect to a network, unless it is your first time powering on, in which case it will tell you it does not have any credentials. This is in preparation for our Web Application which will launch early 2025.

When the device has finished booting, you will land on the [Main Menu](#).

4.2. USB drive

Data is saved to the USB drive as standard and can be analysed when removed from the device. The supplied USB drives are in FAT32 format, if using an alternative USB drive, you must ensure that it is formatted in the same way.

The device can be operated without the USB drive in place. If the USB drive is not in place on device start-up or at the start of an experiment, an error message will show up on the screen and you will be asked if you want to carry out the experiment without the USB. Select 'Yes' or 'No' using the twist/push button. Experiments **will not** save data, and Dissolved Oxygen calibrations **cannot** be started.

4.3. Culture volume

The recommended volume to use in each flask is 15 ml. This can be increased to 20 mL if necessary. **N.B. Note that using less than 15 mL will have serious negative effects on the reliability of all device measurements.** The Liquid Control Module will set the volume to 15 mL regardless of starting volume by nature of the height of the output needle.

4.4. Cleaning the device and its modules

It is recommended to clean contaminated bungs and flasks without oxygen spots via autoclaving. Linked below are a set of cleaning protocols for the care of your Bioreactor system:

Bungs:

Bungs should be removed from flasks soon after an experiment. Their needles should be flushed and the whole bung autoclaved after removing the temperature probe. Plugs and needle covers should be removed prior to being autoclaved and sterilized at room temperature. The filters should be disposed of in lab waste. See below for full detail on Bung cleaning:

dx.doi.org/10.17504/protocols.io.yxmvmeq29g3p/v1

N.B. None of the plugs are rated for autoclave. We have found that all except the blue vent plugs can be autoclaved <10 times without damage. The blue vent plugs may not be autoclaved as they will melt. The needle covers are silicon, and can be autoclaved.

Temperature Probe Cleaning:

The temperature probes should be removed from the bungs post experiment and pre-bung cleaning. **Do not autoclave the bungs with temperature probes still in place.** They can be cleaned by wiping with ethanol or preferred cleaning solution.

N.B, When removing the temperature probes from the bung, please only use the metal grip or plastic jack. Avoid pulling on the wire itself.

Liquid Control Tubing:

Liquid control tubing should be disconnected from the bung needles soon after an experiment and cleaned with the Pump Cleaning protocol. They may be autoclaved a maximum of 5 times if necessary. See below for full detail on Liquid Control Tubing cleaning:

dx.doi.org/10.17504/protocols.io.dm6gpz2jdlzp/v1

Oxygen-Spot Flasks:

The oxygen spot flasks should be washed with de-ionised water and rinsed with ethanol. They should be left to dry after rinsing with ethanol, and stored in a dark place for at least 24 hours after cleaning. These flasks may be autoclaved a limited number of times if necessary, and should be stored in a dark place for at least 1 week after autoclaving. See below for full detail on cleaning Oxygen-Spot flasks:

dx.doi.org/10.17504/protocols.io.261ged277v47/v2

pH Probes:

The pH probes should be removed from the bungs soon after an experiment. Take care to ensure the probes are gently handled, as the glass is delicate. They should be rinsed with de-ionised water and ethanol, and soaked in storage solution (3M potassium chloride) in their rubber storage sleeve. Wrap parafilm around the edge of the sleeve to prevent drying. See below for full detail on pH probe cleaning:

dx.doi.org/10.17504/protocols.io.j8nlkoz31v5r/v2

Routine Cleaning

The device should be wiped with a damp cloth for regular cleaning.

If contaminants come into contact with the flask holders, other parts of the reactors, or the base of the device these should be wiped down with a cloth damped with disinfectant. Do not spray directly onto the system. If you suspect liquid ingress into the unit, power off the device immediately and contact help@ogibio.co.uk.

In choosing a method of disinfection be aware that the enclosure is composed of both polypropylene and metal components and a suitable disinfectant should be chosen.

4.5. Device Main Menu

The [Main Menu](#) will change slightly depending on which modules you have installed, however will always have a default set of options.

- **Batch Culture:** Every device has a basic batch culture function. This experiment mode allows you to use the device to automate analysis of a growing cell culture. See Section 7 for information on setting up one of these experiments.
- **Additional Experiments (Optional):** This refers to a selection of menu items that become available when different modules are installed. These experiments are, in order of appearance:
 - Turbidostat ([Section 8](#)),
 - Chemostat ([Section 8](#)), and
 - pH Control ([Sections 8 and 11](#)).
 - Note that these options will not appear if you do not have the relevant modules installed.
- **Quick Measure:** This menu item allows you to take individual measurements of Temperature, OD, Dissolved Oxygen and pH if the relevant modules are installed([Section 5.4.](#)).
- **Settings:** Here is where you can create and view settings presets for convenient device setup ([Section 7.5](#)).
- **Calibration:** All calibration functions are found here. By default, only the OD calibration ([Section 6](#)) is visible. See Section 6 for more information regarding OD calibration. Each module that requires calibration, Dissolved Oxygen and pH modules, has its own menu item within [Calibration](#) that appears when that module is connected. Please see the individual module section in this manual for information on calibrating the respective modules.

- **Pump cleaning:** This menu item allows you to run the pumps in the Liquid Control module to flush cleaning liquid through the tubing as described in Section 7.3.
- **WiFi:** This menu contains the functions to connect the device to a WiFi network and to the OGI Web App. The Web App will launch in Q1 2025, but in the meantime you can use the menu option WiFi > Network connect. to connect the device to an available WiFi network to automatically retrieve date information for experimental set up. The device will look for available networks and display the network names on the LCD. After selecting the desired network, you will be prompted to enter the credentials to connect to the network. Depending on your network security protocol, these can be either a simple password or a username, an optional anonymous identity and a password. Before entering each required credential, you will be prompted to enter its length (i.e. the number of characters you need to enter).
- **Tests:** A sub-menu of the standard tests run on the device to ensure it is operating correctly. They are provided to support remote diagnostics by our support team and to ensure that your device is still running as you need. We may ask for the results of some of these tests should troubleshooting be necessary. Also contained in this menu is the **Factory Reset** function, which  will restore your device to factory defaults. **N.B. WARNING: Factory Reset cannot be undone and will wipe all calibrations from the device. Use with caution.**

4.6. Quick Measure

Quick Measure allows you to take a single measurement from installed analysis modules (currently Fluorescence Quick Measure is unavailable). As default, each device can measure **Temperature** and **OD**. **Note that these measurements are not saved to the USB.**

- **Temperature:** Displays a real-time measurement from each temperature probe. If a probe isn't connected, the device will display a high-magnitude negative number.

- **OD:** You will be asked which OD calibration you would like to use, then to tell the device which reactors you would like to measure in. The device will then warm up the LEDs for 2 minutes and take a measurement, displaying the results on the LCD screen.
- **Dissolved Oxygen:** You will be asked which O2 calibration you would like to use and then the device will display a real-time measurement from each flask with an O2 sensor spot.
- **pH:** You will be asked to select a pH calibration and then the device will display a real-time measurement from each pH probe.

4.7. Keeping the OGI3 Up To Date

New versions of the OGI3 software may be downloaded from our website. The updater requires a Windows PC or laptop and needs PowerShell 7 to be installed. After extracting the downloaded archive, clicking [runme.exe](#) will open the updater interface. Instructions for running the updater can be found in this program, and are also provided in the readme.

Please ensure that adequate time is allowed for the update and verification processes to complete. Partial updates will fail. Ensure that your laptop is connected to the Board Prog. port for updating the main board and the WiFi Prog. port for updating the WiFi card.

The main board and the WiFi board must be kept at versions that are compatible with each other. If the OGI3 detects any incompatibility, it will report this on the LCD screen. To resolve, simply run the updater and update both boards.

N.B. Do not attempt to update via the front USB port. This port is for storage devices only.

5. Calibration of the device

5.1. Optical Density and Calibration

Optical density is not often a well-explained method of determining cell density, and it can be difficult to know whether one can trust a given method of obtaining it.

Conventional spectrophotometers present OD via measurement of light transmitted through a sample of culture in a cuvette. For cells, most light that is not transmitted is scattered (rather than absorbed). You will probably be familiar with a so-called “linear regime” or “region”, when discussing results taken in your spec. The linear relationship between cell number and OD within this region arises from the fact that at low cell densities each photon is either scattered once or transmitted. This ‘single-scattering’ leads to rays exiting the sample at wide angles, which then do not enter the detector of the spectrophotometer. As the cell number increases, it becomes more likely that rays are scattered multiple times (‘multiple-scattering’). The more times a ray is scattered, the more likely it ends up in the detector, and so the detector will see an increasing contribution from scattered light in addition to transmitted light, meaning that the amount of light the detector sees is no longer directly proportional to the number of cells in the sample, and the relationship between OD and cell density is no longer linear.

In the OGI3, we do not measure transmitted light. Rather, we directly measure the light that has been scattered by cells in the sample. The consequence of this is that the relationship between cell number and OD may not be linear at any point in our device, and so **it requires calibration against true (i.e. within the linear region) readings from a spectrophotometer.** In fact, the relationship between the amount of light we collect (expressed as a ‘Device Reading’) and OD nicely follows the power law shown overleaf.

Once readings have been gathered using the method in Section 6.2., the device will fit according to a power law:

$$y = p_0 + p_1 x^{p_2}$$

where in our case y is OD, x is the reading, p_0 is the y-intercept, p_2 is the power, and p_1 is a scaling factor that would be analogous to the gradient when p_1 is close to 1. We find this relationship fits well to our device readings and ODs. If you wish to perform the manual calibration as in [Section 5.2.](#), you may do so with that formula as the fitting function. You will find in the Scripts folder in the software update package a python script that will plot a calibration file from an OD calibration

5.2. Calibrating your Device

Calibration of the device's OD measurements is performed in experimental conditions, (i.e. each calibration has the same cells, media, and temperature as in respective experiments), beginning with a concentrated culture and diluting down throughout the process. The cells should not be growing throughout the calibration, consider using bacteriostatic antibiotics to halt them when desired. Additionally, when measuring alongside the reactor in a spectrophotometer, please ensure that you are comfortably within the linear region of your spectrophotometer, otherwise the calibration will not be trustworthy. Before you begin, you should have a good idea of the ODs you want to calibrate over. **N.B. You must have at least one point at or near 0, otherwise the low-OD readings will be affected. There is a MAX OD of 5 in the calibration process. For the most accurate readings, we recommend calibrating the OD in each reactor with the flask you intend to use in that reactor for experiments.**

1. Prior to calibration, you should prepare one flask of at least 15 mL of cell culture at a known OD no higher than an OD of 5 per reactor to calibrate.
2. Start the calibration from [Calibration > Calibrate ODs](#).
3. You will be prompted for the date, the name and save location for the calibration.
4. You will be asked to select heating for individual reactors, we recommend calibrating at the temperature you expect to experiment at.

5. The motors will briefly turn on and the LEDs will warm up and you will be presented with options: [Finish/A/B/C/D](#), with a number indicating how many readings are left for each reactor. Note that you do not have to use all 6 readings, three readings minimum are required for calibration. However, if it is your first time using the device or in a given condition, **we strongly recommend doing so.**
6. Select the reactor you wish to take a measurement point in and you will see [Stirring...](#) and then [Measuring](#). After the measurement has been taken, take a measurement of the culture in that reactor in your benchtop spec. Once all the measurements for a given OD have been taken you may dilute the culture and re-measure in the spectrophotometer for all flasks. Continue until you have used all 6 measurements or you are satisfied with the breadth of your calibration.
7. When ready, press [Finish](#). You'll be asked to enter the ODs you measured in your benchtop spec corresponding to all of the measurements made in the device.
8. The calibration values will be saved to memory & to file on the USB drive.

5.3. Manual OD Calibration

Sometimes you may need to make adjustments after the calibration is complete, e.g. if you entered an incorrect OD value or if you want to discard or add points afterwards. In such cases one may retrieve the readings and ODs from the saved file in the USB and manually fit a curve to it. **Before carrying out this operation, please ensure you have read and understood Section 5.1. If anything is not clear, please contact help@ogibio.co.uk.**

In [OD Calibration](#) > [Input Cal](#). one may insert the parameters found by an external fitting, rather than relying on the device to perform the operation.

6. Setting up an experiment

6.1. Starting an experiment from the device

1. On booting up [Batch Culture](#) will be the option on the screen, selecting this will begin an experiment. Liquid Control Module experiments (e.g. Turbidostat) will be visible on this menu to the right of Batch Culture if that module is installed.
- Select [Start Experiment](#) to enter the setup menu. This menu is presented in a flat structure, with options being navigable via scrolling and settable via clicking into them:
 - [Date yyyy-mm-dd](#): Clicking into this option allows you to set the date in the format shown. This will be used as the experiment name for the folder and files saved on the USB.
 - [Load settings](#): We allow the user to save presets of settings, they can be loaded here to save time on experimental setup. Refer to [Section 7.5.](#) to explore presets.
 - [Ambient Temp](#): Here you should select the temperature of the environment the device is in. This will be saved to the settings file, and will help the device determine whether it can control temperature in the reactors.
 - [Temp. Control](#): Clicking into this setting will enable you to set temperature targets for each reactor. The maximum temperature you can set is 50°C, and the minimum temperature is 6°C above the ambient temperature. This is due to the heat that is inherently built up as the device runs.
 - [OD Calibration](#): Clicking into this setting allows you to select an OD calibration for your experiment. At this time, you may not start an experiment without an OD calibration. See [Section 6](#) for notes on calibrating OD.
 - [Exp Length](#): Set the length of your experiment. Note that the experiment can be stopped at any time during an experiment and so unless media usage is a concern, this setting is not critical.
 - [Sampling Rate](#): Set the sampling rate in minutes of your experiment. Please note that the device enforces a 3 minute minimum rate to ensure all processes have adequate time to complete.
 - [Stirring Speed](#): Set the stirring speed in RPM of your experiment. The device can stir at speeds between 500 and 5800 RPM.

- **Start:** If no OD calibration has been selected there will be a prompt here to return to that menu item and do so. If one has been selected you may start the experiment. At this point, the device will check the alignment of oxygen flasks if using.

Once the experiment is running, you will have access to a new menu with options to pause or stop the experiment, update the settings, and view the last measurement.

6.2. During an Experiment

When an experiment is running the most recent measurement of each reactor is displayed on the LCD screen. By clicking the Twist/push knob you may cycle through measurement type (e.g. OD -> O2).

In addition to viewing current measurement readings you have a few options for interacting with the device. Note that while the screen displays **Measuring...** interaction is suspended to ensure data integrity.

- **Stop Experiment:** If you wish to stop the experiment, select this option. It will ask if you are sure with a Yes/No response. Upon stopping, the device will exit the experiment mode and the USB may be taken out to review data.
- **Pause Experiment:** If you wish to pause the experiment, for example if you wish to check data without stopping it, you may use this option. The screen will then update to display **Resume Experiment** for when you are ready for the experiment to continue.
 - a. The time that the experiment was paused will be saved to the settings file, as will the time the experiment was restarted.
 - b. While the experiment is paused you will not be able to change any other settings until you resume the experiment.
 - c. While paused, you may remove the USB to access the data. If the USB drive is not replaced when **Resume Experiment** is clicked, you will be warned that it is not present and asked if you wish to continue. You may replace the USB at this time, or you may continue the experiment without it if you so desire.

N.B. if you remove the files from the USB stick and do not replace them, the

⚠ device will generate a new file when it begins saving measurements again.

This new file will not contain headers.

- **Update Settings:** During an experiment you may update the settings in the device. This option opens a reduced version of the settings menu from in [Section 6.1](#), where you may update the desired settings.

6.3. Data format and handling

- The data recorded during the experiment will be saved to the USB drive. Each experiment is contained in folder with the day's date. Multiple experiments on the same day will generate new folders with the experiment number appended. E.g. 240101_01 would be the second experiment performed on 2024/01/01.
- **Files with suffix 'A'.** These are the analysis files. These contain the time each measurement was taken at and the corresponding OD.
- **Files with suffix 'S'.** These record the settings & calibration used for the experiment. They will also record times of pauses, any settings that were changed and other relevant information about any other modules activated in the experiment.
- **Files with suffix 'TM'.** These are records of the flask temperature during the experiment.

6.4. Saving Settings

It is possible to save settings to memory for convenient loading in an experiment. When you enter the settings menu, pre-sets will show on the screen with details of the settings scrolling on the screen. If you select a saved setting, you will be asked if you want to overwrite the settings in this location. If you want to save a new group of settings in a new location, use the twist/push button to highlight a space labelled [Empty](#) and push the twist/push button to input values for [New Settings](#) in this data space.

To select settings, use the twist button to select the desired value and push to confirm selection. The settings that can be saved are:

1. Experiment Length

- Total length of the experiment in hours.
- Default experiment length is 96 hours.
- Minimum experiment length is 1 hour but an experiment can be stopped at any time by selecting 'Stop Experiment'.

2. Sampling Rate

- The period of time between measurements after which the motors stop spinning and an OD measurement is taken.
- Default sample rate is 5 minutes.
- Minimum sample rate is 3 minutes to ensure all measurements execute correctly.

3. Stirring speed

- The speed at which the motors spin, mixing the culture.
- Stirring speed is in rpm and can be changed in increments of 100 RPM.
- The minimum speed is 0 RPM and the maximum is 5800 RPM. Note that there is a step between 0 RPM and 500 RPM as the motors cannot drive anywhere in between those values.
- Default stirring speed is 4000 RPM.
- If you select 0 RPM, the motors will be off for most of the time during the experiment. However, immediately prior to a measurement cycle they will turn on for 10 seconds at 500 RPM to gently resuspend the cells and avoid sedimentation.

Saved settings can be used in an experiment by selecting them during the experiment setup. New settings can be selected in an experiment setup. Experiment length, stirring speed and sampling rate can all be changed during an experiment in the experiment menu by choosing 'Update Settings'.

Accessory Modules

7. Liquid Control Module (P/N OGI3-1011)

The Liquid Control module is a separate unit designed to sit underneath the OGI3 BioReactor. It can be used to hold continuous cultures at a set target OD from either a high dilution or a pre-grown culture as a turbidostat ([Section 7.6](#)), to run an experiment in chemostat conditions by specifying a continual flow rate of media ([Section 7.11](#)), or to keep cultures at a specified pH rate ([Section 10.5](#)).

7.1. Connecting the Liquid Control module

Ensure that the Liquid Control module is connected to the OGI3 BioReactor before switching on the device. The LCD screen on the OGI3 will display **Liquid ctrl ok** if the module is properly connected.




Figure 4: Liquid Control Module

The BioReactor should be placed on top of the Liquid Control module, and the connection cable inserted into the connection port at the rear of the device (Figure 2).

Please ensure that the screws on the connector are tightened to avoid accidental disconnection of the Liquid Control module from the main device.



Figure 5: Liquid Control module in position

7.2. Fitting the tubing

To operate the Liquid Control system tubing must be fitted to each pump. The device will be delivered without tubing in the pumps. Follow the guide below to put tubing into the pumps.

1. The flap on the pump cover should be sitting parallel to the bench in the locked position (Figure 6). The flaps **must** be in the locked position during a liquid-controlled experiment.



Figure 6: Pump cover flap in the locked position



Figure 7: Pump cover flap in the open position



Figure 8: Line up latch and groove to reconnect pump cover

2. To remove the pump cover, press down on the flap until it is in line with the side of the Liquid Control module (Figure 7). The pump cover can then be removed from the pump so that the tubing can be put in place in the pump cover.
3. Each pump has 2 tubing clips (highlighted in Figure 9, left). The clips are slotted into place in the pump cover with the opening facing out. The tubing should be pressed into the right-hand clip (Figure 9, middle) and wrapped around the tubing guide (Figure 9, middle) and then the second clip slotted in place (Figure 9, right). These clips are necessary for correct pump functioning. Position the tubing so that it can connect without stretching to the appropriate flasks. **Note that as the tubing stiffens with use, you may wish to mark the position of the tube-clamps on the tubing with indelible marker.**




Figure 9: The tubing fits into the pump tube clips which then attach to the pump cover

4. To put the pump cover back on, there are 2 latches that should be slotted into the grooves as shown in Figure 8 and then the flap should be returned to the locked position (Figure 6). Push the cover towards the pump if the movement is stiff.
5. The tubing will naturally stiffen over the first 10 hours of use or so. If you notice that the flow rate drops significantly or becomes unreliable check the tubing. If it is noticeably stiff or crumbly it is time to replace that tubing.

7.3. Cleaning the pump tubing

- The recommended method for cleaning the tubing is with 70% ethanol and 18 MΩ-cm water for rinsing. The tubing can be autoclaved if necessary, however should be replaced after no more than 5 autoclave cycles. The cleaning process utilises the function of the pumps themselves to propagate the cleaning fluid through the length of the tubing. Tubing can also be removed from the pumps and cleaning media flushed through by hand. During the cleaning process, you will have control over how long the cleaning fluid is inside the tubing itself.
- It is recommended to clean the pumps before and after a liquid-controlled experiment has been performed.

N.B. If a liquid-controlled experiment is carried out for a period of several days then note that the risk of tube obstruction is increased. This is dependent on the type of organism being used. Obstructions can affect the function of the pumps and the ability of the device to maintain desired conditions. If obstructions are noticed it is recommended to use a fresh set of tubing for subsequent tests.

In order to clean your Liquid Control tubing, please follow the following instructions carefully.

1. Make up a flask with around 5-10 ml of 70% ethanol. Make up another flask with around 10 ml of distilled water for flushing the pump tubing. (These volumes are recommended for the default tubing length of 85 cm with the minimum period of cleaning recommended – you will require larger volumes for increased cleaning times). Use another empty flask to collect the used fluid.
2. Disconnect tubing from any needles. Place all the inputs (right side of the pumps) in the ethanol, and all the outputs (left side of the pumps) in the empty waste flask. Note that the handedness is the same for all pumps and so one side of the device will be oppositely orientated to the other.
3. Use the main menu to select [Pump Cleaning](#).
4. Then select [Start Cleaning](#).
5. The screen will show [Press button to clean pumps](#).
6. Press the button and the pumps will be turn on and propagate the liquid through the pumps.

7. The screen will show [Press button to stop pumping](#).
8. Once you feel that you have pumped the cleaning fluid through the tubing for long enough (at default tubing length we recommend a minimum of 30s but optimise for your organism), press the button again to turn off and the pumps will turn off.
9. The screen will display [Press button to flush pumps](#).
10. Swap your cleaning fluid for flushing fluid (distilled water)
11. Press the button and the pumps will restart, flushing distilled water through the pumps.
12. The screen will display [Remove tubing + press button](#),
13. Once you feel that you have pumped the flushing fluid through the tubing for long enough (at default tubing length we recommend a minimum of 60s – double the cleaning time), then press the button.
14. The pumps will continue to run while all of the flushing fluid is removed from the tubing and the screen will show [Press button to stop pumping](#).
15. Once you can see that there is no more fluid coming out of the output tubing into the collection flask, press the button and the pumps will turn off. You will then be notified that the cleaning process is complete, and the system will return to the main menu.

N.B. When using brand new tubing, it is recommended that you increase the recommended cleaning times by at least 5 seconds, note that this will require additional cleaning fluid.




Figure 10: Base Bioreactor with a Liquid Control Model fully connected to Reactor C

7.4. Preparing for a liquid-controlled experiment

Before conducting a turbidostat experiment it is recommended to clean the pump tubing. See [Section 7.3.](#) for further information.

1. Set out your source flasks and drain flasks (Figure 11) and culture flasks (Figure 12) for each reactor that you wish to use. 500 ml standard laboratory flasks should be used for the source and drain flasks. The flasks provided should be used as the culture flasks.



Figure 11: Source/Drain flask with bung




Figure 12: Ensure the input flask is connected to S (Source) and the output or waste flask is connected to D (Drain).

2. Use volumes of media that are suitable based on cell growth rates and desired experiment length – refer to [Section 7.6](#). for information on estimating the media usage of your Turbidostat experiment.
3. Put your 15 ml cell cultures into the reactors you wish to use during the experiment.
 - a. The fixed needles in the culture flask bungs should give a maintained culture volume of 15 ml (slight variation may occur due to manufacture).
4. Connect your pump tubing to the appropriate part of the system.
 - a. Refer to [Section 7.2](#). for information on putting the tubing into the pumps.
 - b. **N.B. when connecting your tubing/needles, first wipe with ethanol to remove any contamination on the tubing/needle openings.**
 - c. The source flask bungs each have a long needle (Figure 11) and the drain flask bungs each have a short needle (Figure 12).
 - d. The culture flasks each have a long needle (Drain) and a short needle (Source) (Figure 13).
 - e. Connect the tubing from the source flask bung needle to the culture flask addition needle and from the culture flask removal needle to drain flasks

according to the diagram in Figure 14. Refer to Bioreactor labelling for input and output pumps.

- f. Do this for all reactors that you are using.




Figure 13: Input and output pump configuration showing tubing connections

7.5. Turbidostat Experiment Settings

Section 6.1. covers the basic Batch Culture settings. By connecting a Liquid Control module and starting a Turbidostat experiment you unlock two new sets of settings under special menu items. For [Chemostat](#), skip ahead to [Section 7.9](#).

- **Turbidostat:**

- a. **Maintain ODs:** Enabling this setting will switch all reactors to 'Maintain' mode, and so they will work to maintain a cell culture at the OD it was at when it was placed in the reactor. While this affects all reactors, you can still turn individual reactors on and off by scrolling to the right.

Note that the OD of your culture will be rounded to the nearest 0.05.

- b. If maintain ODs is not selected each reactors target OD can be set individually. To do this select the reactor in this sub-menu, choose whether it is ON or OFF, and set the target OD you desire. You can change this setting later, while the device is running.

- c. **Tube Length:** This is the length of your tubing, and is required to ensure  correct priming function of the Liquid Control module. The default value is set at 85cm - the length of tubing supplied by us. **Please note that [Factory Reset](#) will restore this value to this default value.**

N.B. Please ensure all tubing used is the same length.

7.6. Turbidostat Media Usage Prediction

When setting up a Turbidostat experiment 500 ml bottles should be setup and your fresh media added. In the Turbidostat menu there is a tool for predicting how long it will take for this media volume to be depleted due to dilution of your cultures.

- Input the values for volume of media used in the setup, target OD, doubling time of your culture, culture volume, stock OD, dilution of this stock culture and lag time.
- This will then display an approximate time period in which it is expected that your media will run out for this specific experiment.
- This will display on the screen for 5 seconds before going back to the Turbidostat menu.
- This value can then be used as a guide as to when the experiment must be stopped or fresh media added to the setup. If any of these variables are unknown it is recommended to monitor your first experiment to check the time for use of media.

7.7. During a Liquid Control experiment

After pressing **Start** during the experimental setup, you will be asked if you wish to prime the pumps, i.e. fill the tubing with input liquid up to the input needle. Currently, the slight variation in pump rates means that they will over-prime (and a few drops of fluid will exit the input tube during the process). If you wish to prime accurately, you may at this point disconnect the input tube and prime into a sterile tissue or bottle before replacing the tube. Once priming is complete, you will be able to **Press to Continue** and start the experiment.

N.B. The experiment cannot be stopped during a dilution period, measurement or priming period. The same rules apply to Turbidostat experiments as apply to Batch Culture experiments with regards to USB removal and experiment pausing and stoppage.

N.B. If one of the reactors runs out of fresh media during a turbidostat run, the system will detect that it is not diluting and will turn off this reactor. An error note

will be saved in the settings file. If you then top up the media, the reactor will still remain off (no dilution) so if you require more than 500 ml of media in one experiment, ensure you swap in fresh media before the system has run out and switched off dilution. If turbidostat function on a reactor has been turned off during an experiment, it can be reactivated using the twist/push button and selecting 'Turb. Reactors'.

7.8. Data format and handling

If performing a turbidostat test, then an additional file with suffix 'T' will be generated. This contains the time, OD, the status of each pump, and the difference between measured and target OD.

N.B. If you want to remove the USB drive during an experiment to view your files, first pause the experiment and then remove the drive.

7.9. Chemostat Functionality

The Liquid Control module may also be used as a Chemostat, where experimental conditions are maintained through continuous pumping of fresh media at a given rate. The special settings menu item for this is called [Chemostat](#).

- [Chemostat](#): Each reactor (or all at once) may be given its own unique flow rate. The minimum flow rate you can specify is 0.1 mL/hour, and the maximum is 100 mL/hour. Please ensure that you have enough media for a given flow rate.

7.10. pH Control

New with the OGI3 is the ability to control the pH of your culture. This is achieved through a combination of the Liquid Control module and the pH module, and instructions on carrying out this experiment mode are found in [Section 10](#).

8. Oxygen Module (P/N OGI3-1008)

8.1. Connecting the oxygen module

1. Loosen the screw on the blank cover for the O₂ sector on the side of the flask holder.
2. Remove the blank cover and store the cover and screw for future use.
3. Insert the tab at the bottom of the oxygen module into the hole on the Bioreactor box and plug in the connector at the base of the module.
4. Move the top part of the module's cover to bring it into contact with the flask holder with a rotatory motion.
5. Secure the module to the flask holder with the previously removed screw.

N.B. Screws should be touch-tight. Do not overtighten as this will distort the plastic cover.

8.2. Calibrating the Oxygen Modules

1. **Prepare your 100% oxygen solution.** You will need 15 ml of liquid per flask. For best results, prepare your calibration media in advance and pre-equilibrate it as follows: put the media in a clean bottle or flask, loosen the cap to allow for gas exchange with the environment and let the media equilibrate with air for 1-2h, by shaking at moderate speed and pre-warming the media to the temperature that you are planning to use for your experiments.
2. **Transfer solution into the Bioreactor flasks.** Once the media is warmed up, it can be transferred into flasks with oxygen sensor spots. Insert the flasks into the reactors using the notch in the flask holder to align the oxygen spots with the oxygen detector as shown in Figure 15.
3. **Prepare your 0% oxygen solution.** You will need at least 15 ml of calibration solution per flask. You can either use commercially available 0% oxygen solutions/tablets (please follow manufacturer's instructions for correct use of these), or deplete the oxygen in the media by growing cells in sealed flasks. The 0% oxygen solution should be pre-warmed to the working temperature in a sealed container before starting the calibration.



Figure 15: Correct alignment of oxygen sensor spot – Spot to notch in flask holder

4. **Start calibration.** In the [Calibration](#) menu, select [Calibrate Oxygen](#) and enter the date. You will be prompted to choose a slot. If the slot is empty (i.e. all flasks show Factory) or you do not want to keep any existing data, then select [Overwrite](#) and continue. If the slot has existing calibration data you would like to keep then select [Update](#) and continue. The device will scan for sensors and indicate which reactors have the sensing electronics for dissolved oxygen. **N.B.** The calibration does not need to be performed in a single step.

- 5. Calibrate 100%.** If you want to calibrate 100%, select **Yes** when the device reads **Calibrate 100%**. If there isn't already a USB drive in the device, you will be prompted to insert one now. Then, you will be asked to make sure that the flasks with the calibration solution are in place and to press the button to start the calibration. At this point, the device will check the alignment of the sensor spots and you will be prompted to either adjust the alignment of misaligned sensors or disable individual oxygen modules. If you want to skip the 100% calibration step and proceed to the 0% calibration, you can select **Calibrate 100% > No**. **N.B. It is highly recommended to have both a 100% and a 0% calibration point before running experiments.**
- 6. Wait until the calibration is done.** The Bioreactor will now stir the solution and take oxygen readings every 15s until the measurements become stable. This can take up to 1h but will complete more quickly if your media is pre-equilibrated. The LCD will display letters (**A, B, C, D**) for the flasks that have already completed the calibration as an indication of progress. Once the calibration is complete, you will be asked if you want to save this calibration in the memory of the device. At this point, you can take the USB drive out and check the data saved in the file with suffix 'OH'. If the values settle somewhere around 20 and are consistent they are good, and you can replace the USB drive in the device and select 'Yes'. Note that the settle point may be different between devices.
- 7. Proceed to 0% calibration.** If you want to proceed with the 0% oxygen calibration now, you can now select **Calibrate 0% > Yes**. If there is not already a USB drive in the device, you will be prompted to insert one now. If you want to calibrate the 0% oxygen level at another time, you can repeat the above Bioreactor's setup, skip the 100% calibration, and proceed to the 0% calibration.
- 8. Start 0% calibration.** Once your calibration solution is warmed-up, insert the flasks in the reactor flask holders, aligning the oxygen spots with the oxygen detector. Make sure that the caps are fully sealed. When prompted, press the button to start the 0% calibration process. At this point, the device will check the alignment of the sensor spots and you will be prompted to either adjust the alignment of misaligned sensors or disable individual oxygen modules.
- 9. Wait until the calibration is done.** The Bioreactor will now take oxygen readings until the measurements become stable. This can take up to 1h, but complete

more quickly if the calibration solution is prepared correctly (e.g. ensure that the flasks are fully sealed, do not leave a large empty headspace above the liquid in the flask to prevent diffusion of oxygen into the calibration solution). Once the calibration is complete, you will be asked if you want to save this calibration in the memory of the device. At this point, you can take the USB drive out and check the data saved in the file with suffix 'OL'. If the values settle somewhere around 50 and are consistent they are good, and you can replace the USB drive in the device and select 'Yes'. Note that the settle point may be different between devices.

8.3. Measuring dissolved oxygen in experiments

Having an Oxygen module installed in any reactor will unlock a special menu item for it in experimental setups:

- **Measure O₂?** If you select **Yes**, you will be prompted to select an oxygen calibration and then set the salinity of each reactor.

If oxygen measurement is enabled, the device will check alignment after **Start Experiment** is selected. If you have no oxygen flask in a given reactor, simply scroll once and click **Disable** when you are warned that that flask is not aligned. The device will take oxygen measurements before and after the OD readings and save the results in files with suffixes 'O' and 'OE'. At the beginning of the experiment the device will check the alignment of the sensor spots and you will be prompted to either adjust the alignment of misaligned sensors or disable individual oxygen modules.

8.4. Data format and handling

- **File with suffix 'OH'.** The 100% oxygen calibration will produce an output file with suffix 'OH'. This file contains the raw readings of phase shift (dphi) taken during the calibration process and a status code to indicate errors. Status code 0 indicates that no errors occurred during the oxygen measurement. At the end of the file you will find the results of the calibration (dphi: phase shift in degrees, module temperature in °C, ambient pressure in mbar).

- **File with suffix 'OL'.** The 0% oxygen calibration will produce an output file with suffix 'OL'. This file contains the raw readings of phase shift (dphi) taken during the calibration process and a status code to indicate errors. Status code 0 indicates that no errors occurred during the oxygen measurement. At the end of the file you will find the results of the calibration (dphi: phase shift in degrees, module temperature in °C).
- **File with suffix 'O'.** This file contains the results of the oxygen measurements taken during your experiment. For each reactor, you will find a status code to indicate errors, the dissolved oxygen in $\mu\text{mol/l}$, the dissolved oxygen in mbar, the dissolved oxygen in % of air saturation, the temperature in °C within the module, the signal intensity in mV, the ambient light in mV, the ambient air pressure in mbar and the % relative humidity within the module.
- **File with suffix 'OE'.** This file contains the early oxygen measurement. Each measurement has a corresponding timestamp which is different from those in 'O' allowing the oxygen uptake rate of the culture to be calculated.

8.5. Notes

- Flasks with oxygen sensor spots must be stored such that the spot is kept in the dark.
- It is advisable to always use the same flask in each reactor and recalibrate the oxygen module when you use a different flask.
- Flasks with oxygen spots can be sterilised with 70% ethanol. To preserve the performance and lifetime of the sensor spots, do not leave ethanol in the flask for more than 5 minutes. After cleaning with ethanol, let the flask dry completely and wait 24h before using it again.

If necessary, flasks with oxygen spots may be autoclaved a limited number of times. However, the exact safe number of autoclave cycles is unknown and therefore we **do not** recommend that these flasks are cleaned this way. If the flasks are autoclaved they must be left for at least one week and the system should be recalibrated prior to use.

9. Fluorescence Module (P/N OGI3-1014)

9.1. Connecting the Fluorescence Module

1. Loosen the screws on the blank cover for the fluorescence sector (facing the back of the device).
2. Remove the blank cover, keep the screw.
3. Connect the fluorescence module to the device using the cable, and secure the cover ensuring that the tab at the bottom of the cover fits into the slot in the device's shell.
4. Move the top part of the cover towards the reactor in a rotary motion, hinging on the tab/box connection.
5. Secure the module to the reactor with the previously removed screw.

N.B. Do not overtighten the screw as this will distort the plastic cover.

N.B. This module contains a UV led. This has been tested and assessed in operating mode as Risk Group 0. Please ensure module is firmly secured to the reactor with the fixing screw provided BEFORE turning on the device.

9.2. Description of the Fluorescence Module

The standard configuration of the fluorescence module is the following:

- Excitation LEDS (peak wavelength):
 - LED 1: 365 nm
 - LED 2: 590 nm
 - LED 3: 470 nm
- Detection:
 - Sensor 1 (F1): 400-450 nm
 - F2: 450-500 nm
 - F3: 500-550 nm
 - F4: 550-600 nm
 - F5: 600-650 nm
 - F6: 650-700 nm
 - F7: 750-800 nm

- F8: 800-850 nm

9.3. Using the Fluorescence Module

When the device is switched on, it will automatically detect if the fluorescence modules are installed in the reactors. If a fluorescence module is detected in at least one reactor, you will have an additional **Measure Fluo?** option in the experimental startup menus. There are no calibrations associated with the fluorescence module, so it is a simple **Yes/No** toggle to activate the module. At each measurement cycle, measurements will be taken with all combinations of fitted LEDs and available detection bands. The results of these measurements are saved in a file with suffix 'F'.

Additionally, there will be a **Fluorescence** menu item in the main device menu. In this menu, you will find options to change the intensity of the excitation LEDs.

By default, the LEDs are only turned on for a short time for the fluorescence measurement. If you want to keep an LED on for a prolonged period of time (e.g. for optogenetic activation) you can do so by changing the 'LEDx always on' settings in the Fluorescence menu.

LEDs in the fluorescence module can be turned on/off and the intensity can be changed at any time during the experiment. The time and type of change will be recorded in the settings file.

LEDs that are set "always on" will be switched off for short periods of time during measurements to avoid interference with other optical readings.

9.4. Data Format and Handling

File with suffix 'F'. This file contains the results of the fluorescence measurements. For each reactor, you will find measurements taken with every combination of excitation LEDs and detection bands. The column headers in the file contain the relevant information to interpret the data. For example, LED1_F1_A is the light detected from channel F1 in reactor A when the sample is excited with LED 1.

10. pH Module (OGI P/N OGI3-1012)

The pH module is a new addition with the OGI3 BioReactor. It uses four pH probes that fit into the large hole in the bung to accurately measure the pH of your liquid sample. It can be used in tandem with the Liquid Control module to keep a growing culture at a stable pH.

10.1. Connecting the pH Module

1. If you have purchased a pH module alongside your OGI3 BioReactor unit, it will already be connected.
2. If you purchase a base unit without the pH module it will have a plastic square covering the sparging and pH connectors. To connect the pH module, first remove this square and keep safe.
3. Find the connector on the underside of the pH module, and connect to the 8-pin connector as shown in Figure 16:




Figure 16: Top Connectors and screw holes

4. Sit the module flat on the top of the BioReactor unit, this may take some small rearrangement of the connector wire underneath the module.
5. Use the three butterfly screws provided to fix the module to the base unit. Do not overtighten these screws as this may cause deformation in the top of the base unit and the module casing.

10.2. Handling and Connecting the pH Probes

N.B. The glass pH probes are very fragile. When inserting into the silicon bung be very careful to keep pressure along the length of the probe. Any bending motion or force applied perpendicular to the probe can result in its breakage. We do not accept liability for broken pH probes.

- The pH probes will come with a black storage sleeve wrapped in parafilm. This is how they should be stored when not in use.
- Carefully remove the parafilm and storage sleeve.
- Rinse the tip of the pH probe with pure water (at least 18 MOhm), and insert into the large hole in the silicon bung. The probe should enter without much force required, but if it is a little stiff you may use a light twisting motion to encourage fit.
- Once probe is in place, the bung may be inserted into the culture flask.
- The jack should be connected to the device prior to turning the device on.
- When finished with the pH probes, they should be cleaned and stored in accordance with the following protocol:

[dx.doi.org/10.17504/protocols.io.j8nlkoz31v5r/v2](https://doi.org/10.17504/protocols.io.j8nlkoz31v5r/v2)

10.3. Calibrating the pH module

10.3.1. Configuration of pH Buffers for pH Calibration

The pH module is calibrated through the [Calibrations](#) menu. To calibrate the probes, you will need a set of 2 to 5 pH buffer solutions. There are 3 buffers pre-loaded onto the device, but we recommend configuring your own set as the exact values and temperature coefficients of the solutions may differ brand-to-brand.



pH buffers are generally supplied with a temperature lookup table that describes the change in pH of the buffer solution at different temperatures. This table is usually found on the datasheet, on the manufacturer's website or on the label on the bottle or sachet for the solution (example picture on the left).

To configure a new pH buffer, [select pH probes > pH buffers list](#) in the BioReactor menu and then choose an available slot to configure a new pH buffer. You will be instructed to enter a name for the pH buffer and then to input temperatures and corresponding pH values for the buffer. For each

buffer, you can enter up to 14 temperature – pH value pairs. These pairs should be entered with temperature in ascending order.

The pH buffer data will be saved in the internal memory of the device and can be used for any subsequent calibration.

10.3.2. Calibration of the pH Module

Before you start a calibration, please ensure that the pH probes are firmly connected to the pH module and that each probe is labelled as the reactor that it is connected to. This is to avoid mixing them up if you unplug them for cleaning/storage.

1. To calibrate the pH probes, use the twist/push button to select [Calibration > Calib. pH probes](#).
2. You will be instructed to enter the date and to select an available memory slot to save the pH calibration.
3. Then you will be asked to enter the temperature of the pH buffers that you are going to use for the calibration (e.g. the room temperature if you are doing a calibration at room temperature). All the pH buffers should be at the same temperature to ensure good linearity of the calibration data points.

N.B. The device will apply temperature compensation to the pH measurements during experiments, therefore it is possible to calibrate the pH module at room temperature and then use it for experiments at different temperatures.

4. Then you will need to select the number of buffers that you want to use for the calibration, and select the desired buffers from the list of available buffers in the device memory.
5. Then remove the rubber sleeve from the tip of the probes, rinse the probes with deionised water and dab them dry.
6. Dip the tips of the 4 probes in the buffer solution displayed on the LCD and press the button to start the calibration. The device will now take measurements until the readings are stable and then instruct you to repeat the process with the next buffer solutions.
7. Please remember to rinse the probes with deionised water and dab them dry after each step (i.e. before you dip them in the next buffer solution).
8. When the calibration is complete, rinse the probes with deionised water, dab them dry.
9. Fill the rubber sleeves with fresh storage solution (or 3M KCl), slide the sleeves onto the electrodes' tips and wrap laboratory film around the sleeves to prevent evaporation.

10.4. Measure pH in Experiments

pH can be measured in any experiment mode (e.g. Batch Culture, Turbidostat etc.). When you start an experiment, after setting the main parameters for the culture, you will be asked whether you want to measure pH in the experiment. If you select **Yes**, you will be prompted to select a pH calibration for the measurements.

10.5. pH Control

By utilising both the Liquid Control module and the pH module, you may control the pH level in a sample or culture in the Bioreactor with a single feed with the **pH Control** experiment mode. This mode will control the pH with either an acid **or** a base, but not

both. The pH will be measured according to the sampling time and, if necessary, the source media will be pumped in for 5 seconds. The device will remeasure the pH and repeat if the target pH has not been reached.

The tubing should be cleaned and connected as for the Turbidostat – please see the user manual or quick start guide for instructions. Your source flask should contain the acid or base to be used to correct the pH.

Follow the on screen instructions as usual. The [Select reactors](#) prompt allows you to enable or disable the control of individual flasks. You can set the target pHs of each enabled flask. Finally, choose whether each source bottle contains acid or base.

10.6. Data Handling

- **File with suffix 'PC'.** The pH calibration will produce an output file with suffix 'PC'. This file contains pH values for the buffer solutions and the corresponding raw voltage readings taken during the calibration process. At the end of the file you will find the slope % of the calibration curve, the mV offset and the R^2 value for each probe. A good quality calibration should have R^2 value close to 1, slope between 95-105% and offset between ± 30 mV.
- **File with suffix 'P'.** This file contains the results of the pH measurements taken during your experiment.

Disposal

This product should be sorted for environmental-friendly recycling



Do not dispose of products into household waste!

Only for EU Countries:

According to the European Directive 2012/19/EU on Waste Electrical and Electronic Equipment and its implementation into national law, products that are no longer useable must be collected separately and disposed of in an environmentally friendly manner. At the end of its useful life, the OGI3 BioReactor and any modules may be returned to OGI Bio for disposal if required

Appendix A - Technical Specification

- This device is rated for the power supply provided
 - Power supply input
 - Input voltage: 100-240V AC (+/- 10%)
 - Input current: 1.2-0.5A
 - Input frequency: 50/60Hz
 - Power supply output
 - 15V DC, 6.0A maximum
 - Bioreactor power input
 - 15V DC, 6.0A, 90W
- Fuse: 5A Fast Blow
- The device is intended for indoor use only.
- Pollution Degree: 2
- Operating (ambient) temperature: 15 to 60°C
 - If using the Turbidostat module, maximum operating temperature: 40°C
 - Can be used up to 45°C but this may affect lifetime of pumps.
 - Note that temperature control will still extend to 50°C.
- Relative humidity: 30% - 70%
- Maximum operating altitude: 2000 meters

Appendix B: Troubleshooting

As much as possible, we work to improve your experience of the device and ensure there are no hitches or problems. This isn't always possible however, and so this section aims to provide you with the tools to understand various errors the device may communicate to you. If anything is not clear or your problem persists, please don't hesitate to contact us at help@ogibio.co.uk.

Calibration Errors and Warnings

The device will attempt to ascertain whether or not a calibration has gone well. It may, upon saving or viewing a calibration, give the following messages:

- **Note: <Reactor> Not Calibrated:** This message shows when the reactor detects zero in each of the parameters. If you did not intend to calibrate the reactor it mentions, disregard the message.
- **Warning: <Reactor> Poor Fit:** This indicates that the fit that the device has for the given reactor has a low R^2 value, i.e. that the fit does not accurately represent the OD and reading data. We consider “low” to be below 0.98.
- **Error: <Reactor> Contains Zero:** This is a different error to above, it indicates that a single parameter in the fit is zero. This warrants inspection of the calibration as it is generally a reason to re-calibrate.
- **Error: <Reactor> Intercept -ve:** The ‘-ve’ standing for negative, this indicates that the calibration is showing a negative reading for 0 OD. This is a problem as it means the device will not be able to measure low ODs with any accuracy. In this case, it is generally sufficient to recalibrate the device with a lower OD range.
- **Error: <Reactor> Bad Power:** This indicates that the power parameter, p_2 , is greater than 1. This warrants inspection, and generally requires recalibration.

Oxygen Spot Errors

The Dissolved Oxygen module will provide some error codes in the Oxygen Calibration files if there are any problems, printed in the 'Status' column. Each error code in this column is a cumulative sum of warning and error values from the module. We appreciate that this does add a level of complexity to troubleshooting.

- **Add 1:** Warning – automatic amplification level active
- **Add 2:** Warning – sensor signal intensity low
- **Add 4:** ERROR – optical detector saturated
- **Add 8:** Warning – reference signal intensity low
- **Add 16:** ERROR – reference signal too high
- **Add 64:** Warning – 1000xOxygen enabled
- **Add 128:** Warning – high humidity (>90%RH) within the module
- **Add 256:** ERROR – failure of case temperature sensor
- **Add 512:** ERROR – failure of pressure sensor
- **Add 1024:** ERROR – failure of humidity sensor